

# Cleaning Log

## Maxwell® RSC 48 Instrument (Cat.# AS8500)

Clean the instrument as described in the Maxwell® RSC 48 Instrument Operating Manual TM#510 following any spill or at regular intervals as appropriate for instrument usage. Record activities below.

Serial Number:

**Today's Date** (MM/DD/YYYY): \_\_\_\_\_ **Time:** \_\_\_\_\_

User Name: \_\_\_\_\_

Please check off steps as they are completed:

- Power Off the Instrument
- Disconnect from the Power Source
- Remove Deck Trays
- Clean Magnetic Rod Assembly
- Clean Magnetic Bars
- Clean Plunger Bars
- Clean Deck
- Clean Interior
- Clean Exterior and Door
- Reinsert Deck Tray
- Reconnect to Power Source

Comments: \_\_\_\_\_

---

---

---

---

---

---

---

---

---

---

---

**Today's Date** (MM/DD/YYYY): \_\_\_\_\_ **Time:** \_\_\_\_\_

User Name: \_\_\_\_\_

Please check off steps as they are completed:

- Power Off the Instrument
- Disconnect from the Power Source
- Remove Deck Trays
- Clean Magnetic Rod Assembly
- Clean Magnetic Bars
- Clean Plunger Bars
- Clean Deck
- Clean Interior
- Clean Exterior and Door
- Reinsert Deck Tray
- Reconnect to Power Source

Comments: \_\_\_\_\_

---

---

---

---

---

---

---

---

---

---

---