



Cleaning Log

Maxwell® CSC Instrument (Cat.# AS6000)

Serial Number: _____

Clean the instrument as described in the *Maxwell® CSC Instrument Operating Manual #TM457* following any spill or at regular intervals as appropriate for instrument usage. Record activities below.

Today's Date (MM/DD/YYYY): _____ **Time:** _____

User Name: _____

Please check off steps as they are completed:

- Disconnect from the Power Source.
- Remove Deck Tray.
- Clean Magnetic Rod Assembly.
- Clean Magnet Bar.
- Clean Plunger Bar.
- Clean Deck.
- Clean Interior.
- Clean Exterior and Door.
- Reinsert Deck Tray.
- Reconnect to Power Source.

Comments: _____

Today's Date (MM/DD/YYYY): _____ **Time:** _____

User Name: _____

Please check off steps as they are completed:

- Disconnect from the Power Source.
- Remove Deck Tray.
- Clean Magnetic Rod Assembly.
- Clean Magnet Bar.
- Clean Plunger Bar.
- Clean Deck.
- Clean Interior.
- Clean Exterior and Door.
- Reinsert Deck Tray.
- Reconnect to Power Source.

Comments: _____
